

## GOTECH "the best partner" to raise your competition

### GT-7012-M2 手提式肘節止滑測試儀 Slip Tester (Articulated Strut)

GB/T 20991-2007 · ISO 20344 · ASTM-F1677

本儀器以荷重水準分力與垂直分力比值之原理，測取試件與地面之最大靜摩擦係數。本測試儀為可攜式並可任意於乾、濕或受汙之各種材質地面，測試鞋底、鞋跟材料之抗滑性。

This tester, the portable inclinable articulated strut slip tester, is used to determine the slip resistance of footwear sole, heel or related materials against planar walkway surface or walkway surrogates under dry, wet or contaminated conditions.

|                              |                      |
|------------------------------|----------------------|
| 荷重                           | 10 lb                |
| 試樣面積 Specimen Size           | 3"×3" (76.2×76.2 mm) |
| 測試面積 Test Surface            | 4"×4" 以上 above 4"×4" |
| 底板厚度 Thickness of Main Frame | 3/4"                 |
| 落下高度                         | 1/8"~1/4"            |
| 體積 Dimension (W×D×H)         | 40×22×39 cm          |
| 重量(約) Weight (approx.)       | 15 kg                |



瞭解更多產品信息，請登錄我司網站 [www.gotech.biz](http://www.gotech.biz)